



TAMIL NADU POLLUTION CONTROL BOARD

Circular Memo No.TNPCB / T6 / STP / 05656 / 2016 Dated 01.03.2022

Sub: TNPCB – Monitoring of Sewage Treatment Plants – Calibration of OCEMS and Electromagnetic flow meters – Instructions – Issued - Reg.

- Ref: 1. CPCB File No.A-19014/43/06-UPC-I 4648 dated 7-8-2020 addressed to the Chairman, TNPCB, Chennai.
2. Board's Circular Memo No. TNPCB / T6 / STP / 05656 / 2016 Dated 02.02.2022

All the District Environmental Engineers are informed that CPCB has emphasized that Continuous monitoring is essential to regulate and minimize the inspection of STPs on routine basis and also to ensure compliance of prescribed standards and transmit data to SPCB/PCCs and CPCB on continuous basis.

Accordingly, The CPCB has issued directions to the TNPCB among other things for the installation of Online Continuous Effluent Monitoring System (OCEMS) for self surveillance of STPs in the reference first cited.

Hence, in order to comply with the CPCB's directions instructions were issued vide Board's Circular Memo second Cited.

Further, in order to ensure the importance and accuracy of calibrations of OCEMS and the Electromagnetic flow meters further instructions are issued as follows:

"All the DEEs are instructed to ensure that the calibrations of OCEMS / Electromagnetic flow meters are done only through NABL accredited calibration Laboratories / Approved Laboratories of Weights and Measures Department".


To

All DEEs

Copy to

All JCEEs (M) for necessary follow up

//forwarded by order//


01/03/2022
for Member Secretary

76, MOUNT SALAI, GUINDY, CHENNAI - 600 032.

Tel : 22353134, 22353135, 22353136, 22353137, 22353138, 22353139, 22353140, 22353141

Fax : 044-22353068

Email : tnpceb@md3.vsnl.net.in Web : www.tnpceb.gov.in